

**PENNSYLVANIA DEPARTMENT OF HEALTH
BUREAU OF EMERGENCY MEDICAL SERVICES**

Vehicle #

QRS Inspection Checklist

I. GENERAL INFORMATION:

Date Stickers: Yes No
Decals: Yes No

Name of QRS Agency:

Address:

(Primary Headquarters)

City

State

Zip

License Plate # :

Year:

Make:

Model:

Vehicle Identification # (VIN):

Date Inspected:

Affiliate # :

Regional EMS Council:

Mileage:

	YES	NO	N/A
Was a deficiency notification issued for this vehicle?			
Is a copy of the deficiency notification attached to this form?			
Is a reinspection required?			
VEHICLE/EQUIPMENT	PRESENT AND OPERATING	DEFICIENT	CORRECTED
If vehicle - meets PA Vehicle Code 75 PA C.S.			
Current Vehicle Inspection			
Current Vehicle Insurance			
Current Vehicle Registration			
No Smoking /Oxygen Equipped Sign (1) in front			
Fasten Seat Belts Sign (1) in front			
Interior Requirements:			
Storage Cabinets for equipment or otherwise secured			
Bulky Items Secured			
Radio Equipment (meets regional comm. requirements) Comm. with PSAP and responding ambulances			
MEDICAL SUPPLIES/EQUIPMENT	PRESENT AND OPERATING	DEFICIENT	CORRECTED
Current Version Statewide EMS Protocols			
Portable Suction Unit with wide bore tubing (1) (achieves 300mm/Hg in 4 sec.)			
Suction Catheters: (2 Pharyngeal -Sterile)			
Airways:			
Oropharyngeal (6 different sizes - to include one 0-1,one 2-3, & one 4-5)			
Nasopharyngeal to include one 16-24 Fr. & one 26-34 Fr.) (5 different sizes with lubricating jelly			
MEDICAL SUPPLIES/EQUIPMENT	PRESENT AND OPERATING	DEFICIENT	CORRECTED
Portable O2 cylinder with capacity of at least 300 Liters, with 500 psi Yoke Cylinder with minimum total pressure of 500 psi.			
Gauge/flow meter not gravity dependent and can deliver 0-25 liter per minute.			
Non-sparking wrench/tank opening device			
Secured in vehicle at all times			
Spare O2 Cylinder (Full-300L) (1)			

Oxygen Delivery Devices:			
Nasal Cannulas Adult - Pedi (1 each)			
High Concentration Masks (Adult, Infant and Pediatric - 1 each)			
Pocket Mask with One-Way Valve & O2 port (1)			
Bag Valve Mask Devices-(1)Adult & (1)Pedi (450-700cc) with Adult & Pedi mask			
Sphygmomanometer			
(Child, Adult & Thigh(Lg)-1 each or interchangeable cuffs)			
Stethoscope (Adult & Pediatric - 1 each)			
Penlight (1)			
Dressings:			
Multi-Trauma (10" x 30") (2)			
Occlusive (3" x 4") (4)			
Sterile Gauze Pads (3" x 3") (12)			
Soft Self Adhering (4 rolls)			
Adhesive Tape (4 rolls assort., 1 must be hypoallergenic)			
Bandage Shears (1)			
Regional Approved Triage Tags (20)			
Commercial "Tactical" Tourniquet (1)			
Triangular Bandages (4)			
Sterile Burn Sheets (4' x 4') (2)			
Blankets (2)			
Emergency BLS Jump Kit (1)			
Instant Glucose (45 grams-40% dextrose-d-glucose gel)			
AED (with 2 sets of pads) at least (1) set of pads need to be adult			
	PRESENT AND OPERATING	DEFICIENT	CORRECTED
PERSONAL PROTECTIVE EQUIPMENT			
High-visibility safety apparel (1 per crew member)			
Helmet (1 per crew member)			
Eye Protection -Goggles (1 pair per crew member)			
Work Gloves (1 pair per crew member)			
DOT Emergency Response Guide (1) - Current Ed.			
	PRESENT	DEFICIENT	CORRECTED
VII. PERSONAL PROTECTION EQUIPMENT			
Clear Eye Protection, disposable (1 per crew member)			
Face Mask (N-95 one per crew member)			
Gloves Patient Exam/ Treatment			
Red Bags			
Hand Disinfectant - Non-water (1 container)			
	Inspected By:	(Printed Name)	
	Signature:		
	Date Forwarded to BEMS:		
Electronic Deficiency Form Completed	Yes	N/A	
Digital Images Captured	Yes	N/A	
Vehicle Placed Out of Service (Per I.B. 2013-001)	Yes	N/A	